

Device Schedule:

Intended Purpose as per the Instructions for Use:

The **cobas**® SARS-CoV-2 & Influenza A/B nucleic acid test for use on the **cobas**® Liat® System (**cobas**® SARS-CoV-2 & Influenza A/B) is an automated multiplex real-time RT-PCR assay intended for the simultaneous rapid in vitro qualitative detection and differentiation of SARS-CoV-2, influenza A, and influenza B virus RNA in healthcare provider-collected nasopharyngeal and nasal swabs and self-collected nasal swabs (collected in a healthcare setting with instruction by a healthcare provider) from individuals suspected of a viral respiratory infection.

cobas® SARS-CoV-2 & Influenza A/B is intended for use in the simultaneous rapid in vitro detection and differentiation of SARS-CoV-2, influenza A virus, and influenza B virus nucleic acid in clinical specimens and is not intended to detect influenza C virus. SARS-CoV-2, influenza A and influenza B viral RNA is generally detectable in respiratory specimens during the acute phase of infection. Positive results are indicative of active infection but do not rule out bacterial infection or co-infection with other pathogens not detected by the test. Clinical correlation with patient history and other diagnostic information is necessary to determine patient infection status. The agent detected may not be the definite cause of disease.

Negative results do not preclude infection from SARS-CoV-2, influenza A, and/or influenza B and should not be used as the sole basis for diagnosis, treatment or other patient management decisions. Negative results must be combined with clinical observations, patient history, and/or epidemiological information.

cobas® SARS-CoV-2 & Influenza A/B is intended for use by health professionals or trained operators who are proficient in using the **cobas**® Liat® System in Near Patient Testing, Point of Care (POC), or clinical laboratory settings.

Device Name	Model	Type (Codes as per (EU) 2017/2185)	Risk Classification	Basic UDI-DI
cobas ® SARS-CoV-2 & Influenza A/B Nucleic acid test for use on the cobas® Liat® System	09211101190	IVR 0503	Class D	761333602369B6
cobas ® SARS-CoV-2 & Influenza A/B Quality Control Kit for use on the cobas® Liat® System	09211128190	IVR 0503	Class D	761333602369B6